

ORAL PRESENTATION



Real-time MRI feedback of cavitation ablation therapy (histotripsy)


Steven Allen^{*}, Timothy Hall

From Current and Future Applications of Focused Ultrasound 2014. 4th International Symposium Washington, D.C, USA. 12-16 October 2014

Background/introduction

Histotripsy ablation of liver tumors is a non-invasive surgery that uses high-intensity acoustic pulses to control an inertial cavitation cloud. Repeated exposure to the cavitation cloud renders a target tissue into a homogenous slurry which is then gradually absorbed by the body. Like other non-invasive surgeries, histotripsy requires a feedback system that can estimate therapy location and dose in real-time. Because the time-average power output of histotripsy is very small, the treatment region does not express a significant rise in temperature, making MRI thermometry an ineffective method to actively monitor therapy. Previous work has shown that MRI pulse sequences can be made sensitive to the chaotic water flow present during inertial cavitation. This is done by synchronizing the timing of the histotripsy pulses with the timing of the sequence's gradient waveforms. Incoherent water flow caused by cavitation attenuates the MR signal through a process similar to that used in diffusion-weighted MRI. It was shown that these sequences can give localized contrast specific to histotripsy bubble clouds. However, the MR images could only be acquired in a piecemeal fashion over several minutes such that a single image represented the influence of many cavitation events. Here, we introduce a single-shot MR acquisition sequence that is able to rapidly acquire a complete MR image and remain sensitive to histotripsy cavitation. This is done by synchronizing each histotripsy pulse with incoherent motion-weighting gradients placed just before the readout portion of the sequence. When repeated at the same rate as the histotripsy pulses, the resulting MR images can give feedback on the location of every cavitation cloud applied to the target tissue.

University of Michigan, Ann Arbor, Michigan, United States



Methods


An MR-compatible transducer was placed in the bore of a 7T MRI scanner and coupled to a water bath and a 2 cm surface RF coil. A cavitation sensitive, single-shot, spiral readout sequence was used to repeatedly image a plane transecting the focus of the transducer. Acoustic pulses from the transducer were triggered to fire between the incoherent-motion sensitizing gradients of the sequence and the readout gradients. See figure (1). Control images were also acquired with the sensitizing gradients disabled.

Results and conclusions

An example control image and an example cavitation weighted image are shown in figure 2. In the figure, the surface coil is centered about the transducer focus while the transducer itself and the majority of the water bath lie outside the field of view. The acoustic pulses propagate



© 2015 Allen and Hall; licensee BioMed Central Ltd. This is an Open Access article distributed under the terms of the Creative Commons Attribution License (http://creativecommons.org/licenses/by/4.0), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited. The Creative Commons Public Domain Dedication waiver (http://creativecommons.org/publicdomain/zero/1.0/) applies to the data made available in this article, unless otherwise stated.



from bottom to top of the image. The cavitation weighted image displays a small region of signal attenuation at the transducer focus which is not present in the control image. Each image is acquired over the course of 30 ms and repeated every 400 ms, resulting in localization information for every cavitation pulse applied to the target.

Acknowledgements (Funding)

Author Timothy Hall has equity, consulting, and royalty interests in Histosonics Inc. This work is funded in part by NIH grants R01DK087871, R01CA134579.

Published: 30 June 2015

doi:10.1186/2050-5736-3-S1-O89 Cite this article as: Allen and Hall: Real-time MRI feedback of cavitation ablation therapy (histotripsy). Journal of Therapeutic Ultrasound 2015 3 (Suppl 1):O89.

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at www.biomedcentral.com/submit

BioMed Central